

AS3000 CLINICAL QUICK START GUIDE

AS3000[™] Anesthesia Delivery System

This guide does not replace the AS3000[™] Anesthesia Delivery System and Patient Monitor Operating Instructions. Consult each product's instruction manual for complete setup instructions.

1. **Ensure all flows are OFF and the Y-piece is OPEN prior turning ON the AS3000.**
2. **Follow the ventilator's onscreen instructions** to complete the START UP test.
3. **If the Leak Test fails**, eliminate the leak and retest, or power cycle and bypass the Leak Test to allow the ventilator to operate with the leak.
4. **Turn on the Patient Monitor and Gas Module** 15 minutes prior to the first case of the day and ensure all patient cables and interconnect cables are free from damage and are firmly connected.
5. **Start the ventilator** by using the knob to highlight (white) the MODE tile. Scroll to CMV, PCV, PS or SIMV and press the knob twice.
6. **The STANDBY screen displays no data.** See patient data in MANUAL Mode when using the bag.
7. **When reducing N₂O flow, O₂ decreases proportionally to maintain the same ratio.** Adjust O₂ first and last.
8. **To turn OFF the volume, pressure and Apnea alarms in MANUAL mode** use the knob to select the ALARM Off option.



Assistance is provided by phone, Monday through Friday between 9:00am and 5:00pm EST. Contact your local representative for assistance after hours. Please set aside time between clinical cases when calling for product assistance. Call 1-800-288-2121.

9. The **MANUAL/AUTO** key switches back and forth between the bag and the ventilator.
10. **Confirm the waste gas scavenger** is connected and its (red) float is between minimum and maximum.
11. **ADULT/CHILD** size can only be changed while in the STANDBY mode.
12. **To calibrate the O² sensor**, press MENU. Use the knob to navigate to the CALIBRATE tab (top of screen). Follow on screen instructions exposing sensor to room air for 3 minutes. Always select and enter 'DONE' before returning to the Normal Screen.



Assistance is provided by telephone, Monday through Friday between 9:00am and 5:00pm EST. Contact your local Representative for assistance after hours. Please set aside time between clinical cases when calling for product assistance. Call 1-800-288-2121.